



Republic of the Philippines

DEPARTMENT OF SCIENCE AND TECHNOLOGY

PHILIPPINE SCIENCE HIGH SCHOOL - SOCCSKSARGEN REGION CAMPUS

Bids and Awards Committee

SUPPLEMENTAL/BID BULLETIN NO. 1

**SUPPLY AND DELIVERY OF TECHNICAL AND SCIENTIFIC EQUIPMENT
FOR SCIENCE LABORATORIES**

PhilGEPS Bid Reference No. 10687495

This Supplemental/Bid Bulletin, dated April 3, 2024, is issued to modify, amend, and clarify certain items contained in the **SUPPLY AND DELIVERY OF TECHNICAL AND SCIENTIFIC EQUIPMENT FOR SCIENCE LABORATORIES**.

ITEM NO.	QUERY	CLARIFICATIONS/AMENDMENTS/INCLUSIONS
1	Can Ultralow Temperature Freezer be 100 liters for the storage volume?	As specified in the Technical Specification under General Requirement, Storage Volume is at least 105 liters (3.7 cu.ft.).
2	Can Ultralow Temperature Freezer temperature range be -86°C to -40°C?	As specified in the Technical Specification under General Requirement, Temperature Range is -86 to -20°C at 32°C ambient, hence, temperature of -86 to -40°C is acceptable as it is within range, however, maximum should be at -86°C.
3	Is the bidding per item or per lot?	Bids to be submitted must be by lot.

This Supplemental/Bid Bulletin No. 1 shall form part of the documents under this procurement. Any provisions in the Procurement Documents inconsistent herewith are hereby amended, modified, and superseded accordingly.

For guidance and information of all concerned.

Issued this 3rd day of April 2024 at PSHS-SRC, City of Koronadal, South Cotabato.


JONATHAN A. MADRONERO
BAC Chairperson